

Instruction for Use

diarellaPVL-MRSA

real time PCR Kit

For qualitative *in vitro* detection of Methicillin Resistant Staphylococcus aureus (MRSA) DNA and the differentiation of Community-acquired (CA) and Hospital-acquired (HA) MRSA.

REF

G01120-32

G01120-96



32

96



gerbion GmbH & Co. KG
Remsstr. 1
70806 Kornwestheim
Germany
phone: +49 7154 806 20 0
fax: +49 7154 806 20 29
e-mail: info@gerbion.com
www.gerbion.com

IVD



Index

1	Intended Use.....	3
2	Background Information	3
3	Principle of the Test	4
4	Package Contents.....	4
5	Equipment and Reagents to be Supplied by User.....	5
6	Transport, Storage and Stability	5
7	Warning and Precautions.....	5
8	Sample Preparation	6
9	Control DNA.....	6
10	Real time PCR.....	7
10.1	Important Points Before Starting:	7
10.2	Procedure.....	7
10.3	Instrument Settings.....	9
11	Data Analysis.....	11
12	Assay Validation	13
13	Limitations	13
14	Troubleshooting.....	13
15	Kit Performance	15
15.1	Analytical Sensitivity.....	15
15.2	Linear Range.....	18
15.3	Analytical Specificity.....	19
15.4	Precision	21
16	Abbreviations and Symbols.....	22
17	Literature	22

1 Intended Use

The diarellaPVL-MRSA real time PCR is a multiplex real time PCR for the qualitative detection and differentiation of DNA of methicillin-resistant *Staphylococcus aureus* (MRSA, PVL-negative) and methicillin-sensitive *Staphylococcus aureus* (dropout mutant of MRSA, MS-MRSA, PVL-negative) or methicillin-resistant, coagulase-negative *Staphylococci* (MR-ConS, PVL-negative) and the qualitative detection and differentiation of DNA of methicillin-resistant, community-associated *Staphylococcus aureus* (CA-MRSA, PVL-positive) and methicillin-sensitive, community-associated *Staphylococcus aureus* (CA-MSSA, PVL-positive) or methicillin-resistant coagulase-negative *Staphylococci* (MR-ConS, PVL-positive).

2 Background Information

Staphylococcus aureus are gram-positive coccal bacteria which are ubiquitously found in the environment. About 25-30 % of the human population are long-term carriers of *S. aureus* because the bacteria are frequently part of the skin flora found in the nose and on skin. *S. aureus* can cause a range of illnesses such as minor skin infections, like furuncles and abscesses, pyomyositis, but also life-threatening diseases such as pneumonia, endocarditis, toxic shock syndrome (TSS), and sepsis.

Of increasing importance worldwide are Methicillin-resistant *Staphylococcus aureus* (MRSA) strains. Especially in hospitals MRSA present a danger, because they are resistant to all β -lactam antibiotics (e.g. penicillin) and often possess further resistances to other antibiotics. MRSA is the leading cause of nosocomial infections worldwide (hospital-acquired MRSA also called HA-MRSA). Beside HA-MRSA infections also community-acquired MRSA infections (CA-MRSA) occur, which are acquired outside the hospital. In the recent years also MRSA infections associated with livestock (livestock-associated MRSA or LA-MRSA) emerged, especially with pig farmers.

Since the mid-1990s the number of infections in the population increased with no previously history of medical facility contact. This increase in infections in the population is caused by *Staphylococcus aureus* strains that carry the virulence factor Panton-Valentine leukocidin. Infections tend to occur in healthy younger people. PVL can be produced by methicillin-sensitive MSSA as well as MRSA. MRSA strains that carry the virulence factor PVL are called CA-MRSA. Panton-Valentine leukocidin (PVL) is a bicomponent, poreforming cytotoxin. The cytotoxin of PVL lyses macrophages as well as neutrophil granulocytes and contributes to tissue necrosis. The clinical manifestation of PVL-positive *Staphylococcus aureus* strains are skin and soft tissue infections, particularly recurrent invasive abscesses. Rarely necrotizing pneumonia develops with a mortality rate of up to 75%. Risk groups for transmission CA-MRSA or PVL-MSSA are for example families, persons performing close contact sports, persons from educational settings, prisoners and military personnel.

3 Principle of the Test

The diarellaPVL-MRSA real time PCR contains specific primers and dual-labeled probes for the amplification and detection of *MRSA* DNA in biological specimens. The PCR targets the *SCCmec/orfX* junction and allows for the detection of *MRSA* in biological samples, even those containing Coagulase-Negative Staphylococci. Furthermore, diarellaPVL-MRSA real time PCR Kit allows the detection of the methicillin resistance gene *mecA/mecC*, to eliminate false positive results through dropout mutants. Additionally, the gene of PVL is detected.

The presence of nucleic acid is detected by an increase in fluorescence due to hydrolysis of the probes during amplification. The fluorescence of the pathogen-specific *SCCmec/orfX* probes is measured in the FAM channel. The fluorescence of the *mecA/mecC* gene specific probes is measured in the Cy5 channel. For a positive *MRSA* result, both channels need to show an amplification. The fluorescence of PVL-specific probes is detected in the ROX channel.

Furthermore, diarellaPVL-MRSA real time PCR contains a Control DNA, which is added during DNA extraction and detected in the same reaction by a differently labeled probe.

The Control DNA allows the detection of PCR inhibition and acts as control, that the nucleic acid was isolated from the biological specimen. The fluorescence of the Control DNA is measured in the HEX-channel.

4 Package Contents

The reagents supplied are sufficient for 32 or 96 reactions respectively.

Table 1: Components of the diarellaPVL-MRSA real time PCR.

Label	Lid Colour	Content	
		32	96
Reaction Mix	yellow	1 x 512 µl	1 x 1536 µl
Positive Control	red	1 x 50 µl	1 x 100 µl
Negative Control	green	1 x 50 µl	1 x 100 µl
Control DNA	colourless	1 x 160 µl	1 x 480 µl

5 Equipment and Reagents to be Supplied by User

- DNA isolation kit (e.g. NukEx Pure RNA/DNA, gerbion Cat. No. G05004 or NukEx Mag RNA/DNA, gerbion Cat. No. G05012)
- PCR grade Water
- Sterile microtubes
- Pipets (adjustable volume)
- Sterile pipet tips with filter
- Table centrifuge
- Vortexer
- Real time PCR instrument
- Optical PCR reaction tubes with lid
- Optional: Liquid handling system for automation

6 Transport, Storage and Stability

The diarellaPVL-MRSA real time PCR Kit is shipped on dry ice or cool packs. All components must be stored at maximum -18°C in the dark immediately after receipt. Do not use reagents after the date of expiry printed on the package.

Up to 20 freeze and thaw cycles are possible.

For convenience, opened reagents can be stored at +2-8°C for up to 6 months.

Protect kit components from direct sunlight during the complete test run.

7 Warning and Precautions

Read the Instructions for Use carefully before using the product.

Before first use check the product and its components for:

- Use of this product is limited to personnel specially instructed and trained in the techniques of real-time PCR procedures.
- Specimens should always be treated as infectious and/or biohazardous in accordance with safe laboratory procedures.
- Avoid microbial and nuclease (DNase/RNase) contamination of the eluates and the components of the kit.
- Always use DNase/RNase-free disposable pipette tips with aerosol barriers.
- Always wear protective disposable powder-free gloves when handling kit components.
- Use separated and segregated working areas for (1) sample preparation, (2) reaction setup and (3) amplification/detection activities. The workflow in the laboratory should proceed in unidirectional manner. Always wear disposable gloves in each area and change them before entering a different area.
- Dedicate supplies and equipment to the separate working areas and do not move them from one area to another.
- Store positive and/or potentially positive material separated from all other components of the kit.

- Do not open the reaction tubes/plates post amplification, to avoid contamination with amplicons.
- Additional controls may be tested according to guidelines or requirements of local, state and/or federal regulations or accrediting organizations.
- Do not autoclave reaction tubes after the PCR, since this will not degrade the amplified nucleic acid and will bear the risk to contaminate the laboratory area.
- Discard sample and assay waste according to your local safety regulations.

8 Sample Preparation

Purified DNA is suitable for downstream processing in real time PCR. For the extraction and purification of DNA from various biological materials, commercial kits are available. The operator needs to evaluate the suitability of respective DNA extraction kit.

Commercial kits for RNA isolation such as the following are recommended:

- NukEx Pure RNA/DNA, gerbion
- NukEx Mag RNA/DNA, gerbion

Important:

In addition to the samples always run a ‚water control‘ in your extraction. Treat this water control analogous to a sample.

Comparing the amplification of the Control DNA in the samples to the amplification of the internal control in the water control will give insights on possible inhibitions of the real time PCR. Furthermore, possible contaminations during DNA extraction will be detectable.

Please note the chapter ‚Control DNA‘.

If the real time PCR is not performed immediately, store extracted DNA according to the instructions given by the manufacturer.

9 Control DNA

A Control DNA is supplied and can be used as extraction control or only as inhibition control. This allows the user to control the DNA isolation procedure and to check for possible real time PCR inhibition.

a) Control DNA used as Extraction Control:

diarellaPVL-MRSA real time PCR Control DNA is added to the DNA extraction.

Add 5 µl Control DNA per extraction (5 µl x (N+1)). Mix well. Perform the DNA isolation according to the manufacturer’s instructions. Please follow protocol A.

The Control DNA must be added to the Lysis Buffer of the extraction kit.

b) Control DNA used as Internal Control of the real time PCR:

If only inhibition will be checked please follow protocol B.

10 Real time PCR

10.1 Important Points Before Starting:

- Please pay attention to the chapter ‚Warnings and Precautions‘.
- Before setting up the real time PCR familiarise yourself with the real time PCR instrument and read the user manual supplied with the instrument.
- The programming of the thermal profile should take place before the PCR set up.
- In every PCR run one Positive Control and one Negative Control should be included.
- Before each use, all reagents should be thawed completely at room temperature, thoroughly mixed and centrifuged very briefly.

10.2 Procedure

If the Control DNA is used to control both, the real time PCR and the DNA isolation procedure, please follow protocol A. If the Control DNA is solely used to detect possible inhibition of the real time PCR, please follow protocol B.

Protocol A

The Control DNA was added during DNA extraction (see chapter ‚Control DNA‘). In this case, prepare the Master Mix according to Table 2.

The Master Mix contains all of the components needed for PCR except the sample. Prepare a volume of Master Mix for at least one sample more than required, in order to compensate for pipetting inaccuracy.

Table 2: Preparation of the Master Mix (Control DNA was added during DNA extraction)

Volume per Reaction	Volume Master Mix
16.0 µl Reaction Mix	16.0 µl x (N+1)

Protocol B

The Control DNA is used for the control of the real time PCR only (see chapter ,Control DNA'). In this case, prepare the Master Mix according to Table 3.

The Master Mix contains all of the components needed for real time PCR except the sample. Prepare a volume of Master Mix for at least one sample more than required, in order to compensate for pipetting inaccuracy.

Table 3: Preparation of the Master Mix (Control DNA is added directly to the Master Mix)

Volume per Reaction	Volume Master Mix
16.0 µl Reaction Mix	16.0 µl x (N+1)
0.5 µl Control DNA*	0.5 µl x (N+1)*

*The increase in volume caused by adding the Control DNA is not taken into account when preparing the PCR assay.

Protocol A and B: real time PCR set up

- Place the number of optical PCR reaction tubes needed into the respective tray of the real time PCR instrument.
- Pipet **16 µl** of the Master Mix into each optical PCR reaction tube.
- Add **4 µl** of the eluates from the DNA isolation (including the eluate of the water control), the Positive Control and the Negative Control to the corresponding optical PCR reaction tube (Table 4).
- Close the optical PCR reaction tubes immediately after filling in order to reduce the risk of contamination.

Table 4: Preparation of the real time PCR

Component	Volume
Master Mix	16.0 µl
Sample	4.0 µl
Total Volume	20.0 µl

10.3 Instrument Settings

For the real time PCR use the thermal profile shown in Table 5.

Table 5: real time PCR thermal profile

Description	Time	Temperature	Number of Cycles
Reverse Transcription	10 min	45°C	1
Initial Denaturation	5 min	95°C	1
Amplification of DNA			
Denaturation	10 sec	95°C	45
Annealing	and 40 sec	60°C	
Extension	Acquisition at the end of this step		

The real time PCR thermal profile mentioned represents the universal settings for gerbion real time PCR and real time RT-PCR kits. Therefore, different kits can be used in the same run. For gerbion real time PCR kits used for amplification of DNA, the reverse transcription can be omitted. Dependent on the real time PCR instrument used, further instrument settings have to be adjusted according to the table below.

Table 6: Overview of the instrument settings required for the diarellaPVL-MRSA real time PCR.

Real time PCR Instrument	Parameter	Detection Channel	Notes		
LightCycler 480II			Colour Compensation Kit Multiplex 1 (G070MP1-cc) required		
			Melt Factor	Quant Factor	Max Integration Time (sec)
	SCCmec/orfX	FAM (465-510)	1	10	1
	Control DNA	HEX (533-580)	1	10	2
	PVL mecA/mec	ROX (533-610) CY5 (618-660)	1 1	10 10	2 3
Stratagene Mx3000P / Mx3005P Aria Mx	SCCmec/orfX	FAM	Gain 8	Reference Dye: None	
	Control DNA	HEX	Gain 1		
	PVL	ROX	Gain 1		
	mecA/mecC	Cy5	Gain 4		
ABI 7500 CFX96	SCCmec/orfX	FAM	Option Reference Dye ROX: NO		
	Control DNA	JOE			
	PVL	ROX			
	mecA/mecC	Cy5			
Rotor-Gene Q, Rotor-Gene 3000 Rotor-Gene 6000	SCCmec/orfX	Green	Gain 5		
	Control DNA	Yellow	Gain 5		
	PVL	Orange	Gain 5		
	mecA/mecC	Red	Gain 5		
mic Q-PCR cyclcr	SCCmec/orfX	Green	Gain 8		
	Control DNA	Yellow	Gain 10		
	PVL	Orange	Gain 10		
	mecA/mecC	Red	Gain 10		

11 Data Analysis

Following results can occur:

FAM	ROX	Cy5	HEX	MRSA	
SCCmec/ orfX	PVL	mecA/ mecC	Control DNA		
+	+	+	positive or negative*	positive	Community-acquired MRSA (CA-MRSA, PVL-positive). The result for the Control DNA is irrelevant
+	-	+	positive or negative*	positive	Hospital-acquired MRSA (HA-MRSA, PVL-negative). The result for the Control DNA is irrelevant
+	+	-	positive or negative*	negative	CA-MS-MRSA (methicillin sensitive). The result for the Control DNA is irrelevant
+	-	-	positive or negative*	negative	HA-MS-MRSA. The result for the Control DNA is irrelevant
-	+	+	positive or negative*	negative	CA-MSSA and MR-ConS. The result for the Control DNA is irrelevant
-	+	-	positive or negative*	negative	CA-MSSA. The result for the Control DNA is irrelevant
-	-	+	positive or negative*	negative	MR-ConS. The result for the Control DNA is irrelevant
-	-	-	≤ 34**	negative	MRSA negative
-	-	-	> 34**/ negative	?	Not interpretable

*A strong positive signal in the FAM, Cy5 and/or ROX channel can inhibit the IC. In such cases the result for the Control DNA can be neglected.

**Depending on the PCR instrument and/or the chosen extraction method, the Ct values might be shifted. The water control can be used as reference. If the HEX Ct value of a sample differs a lot from the water control, partial inhibition has occurred, leading to false negative results in case of weak positive samples.

Figure 1 and Figure 2 show examples for positive and negative real time PCR results.




Figure 1: The positive sample shows amplification signal in the specific channel (FAM/Cy5/ROX), whereas no fluorescence signal is detected in the negative sample.




Figure 2: The positive sample as well as the negative sample show a signal in the Control DNA specific HEX channel. The amplification signal of the Control DNA in the negative sample shows, that the missing signal in the specific channel (FAM/Cy5/ROX) is not due to PCR inhibition or failure of DNA isolation, but that the sample is a true negative.

12 Assay Validation

Set a threshold as follows:

Negative Controls

All negative controls should be below the threshold. If there is a potential contamination (appearance of a curve in the negative control or a cluster of curves in specimens at high C_T – for example above 36), results obtained are not interpretable and the whole run (including extraction) has to be repeated.

Positive Controls

All the positive controls must show a positive (i.e. exponential) amplification curve. The positive controls must fall below a C_T of 30.

Internal Controls

All internal controls must show a positive (i.e. exponential) amplification curve. The internal control must fall below a C_T of ≤ 34 . If the internal control is above C_T 34, this points to a purification problem or a strong positive eluate that can inhibit the IC. In the latter case, the assay is valid. If a water control run is performed, the IC must fall below a C_T of ≤ 34 .

13 Limitations

- Strict compliance with the Instructions for Use is required for optimal results.
- Use of this product is limited to personnel specially instructed and trained in the techniques of real-time PCR and in vitro diagnostic procedures.
- Good laboratory practice is essential for proper performance of this assay.
- All reagents should be closely monitored for impurity and contamination. Any suspicious reagents should be discarded.
- This assay must not be used on a biological specimen directly. Appropriate nucleic acid extraction methods have to be conducted prior to using this assay.
- The presence of PCR inhibitors may cause false negative or invalid results.
- As with any diagnostic test, results of the diarellaPVL-MRSA real time PCR kit need to be interpreted in consideration of all clinical and laboratory findings.

14 Troubleshooting

The following troubleshooting guide is included to help you with possible problems that may arise when performing a real time PCR. If you have further questions, please do not hesitate to contact our scientists on info@gerbion.com.

No fluorescence signal in the specific channels of the Positive Control

The selected channel for analysis does not comply with the protocol	Select the channel according to chapter ‚Instrument Settings‘.
Incorrect configuration of the real time PCR	Check your work steps and compare with chapter ‚Procedure‘.
The programming of the thermal profile is incorrect	Compare the thermal profile with chapter ‚Instrument Settings‘.
Incorrect storage conditions for one or more kit components or kit expired	Check the storage conditions and the date of expiry printed on the kit label. If necessary, use a new kit and make sure kit components are stored as described in ‚Transport, Storage and Stability‘.

Weak or no signal of the Control DNA and simultaneous absence of a signal in the specific channels.

real time PCR conditions do not comply with the protocol	Check the real time PCR conditions (chapter ‚Control DNA‘).
real time PCR inhibited	Make sure that you use an appropriate isolation method (see chapter ‚Sample Preparation‘) and follow the manufacturer’s instructions. Make sure that the ethanol-containing washing buffers have been completely removed. An additional centrifugation step at high speed is recommended before elution of the DNA.
DNA loss during isolation process	In case the Control DNA was added before extraction, the lack of an amplification signal can indicate that the DNA isolation was not successful. Make sure that you use an appropriate isolation method (commercial kits are recommended) and stick to the manufacturer’s protocol.
Incorrect storage conditions for one or more components or kit expired	Check the storage conditions and the date of expiry printed on the kit label. If necessary, use a new kit and make sure kit components are stored as described in chapter ‚Transport, Storage and Stability‘.

Detection of a fluorescence in the specific channels of the Negative Control

Contamination during preparation of the PCR	Repeat the real time PCR in replicates. If the result is negative in the repetition, the contamination occurred when the samples were pipetted into the optical PCR reaction tubes. Make sure to pipet the Positive Control last and close the optical PCR reaction tube immediately after adding the sample. If the same result occurs, one or more of the kit components might be contaminated. Make sure that work space and instruments are decontaminated regularly. Use a new kit and repeat the real time PCR.
---------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

15 Kit Performance

15.1 Analytical Sensitivity

The limit of detection (LoD) of diarellaPVL-MRSA real time PCR was determined using serial dilutions of synthetic target DNA-sequences in a Stratagene Mx3005 real time PCR instrument. The LoD of diarellaPVL-MRSA real time PCR for MRSA is at least 10 copies per reaction each.




Sample	Concentration copies/ reaction	C _T -value	mean C _T
		FAM-channel	
SCCmec/ orfX	100000000	12,10	12,10
		11,93	
		12,28	
SCCmec/ orfX	10000000	15,86	15,87
		16,00	
		15,76	
SCCmec/ orfX	1000000	19,40	19,33
		19,26	
		19,32	
SCCmec/ orfX	100000	22,54	22,44
		22,24	
		22,55	
SCCmec/ orfX	10000	25,87	25,82
		25,73	
		25,87	
SCCmec/ orfX	1000	28,60	28,99
		29,14	
		29,22	
SCCmec/ orfX	100	29,88	29,92
		30,37	
		29,52	
SCCmec/ orfX	10	31,31	31,90
		31,90	
		32,49	
SCCmec/ orfX	1,0	32,43	32,12
		32,16	
		31,76	

Sample	Concentration copies/ reaction	Cr-value	mean Cr
		Cy5-channel	
mecA/mecC	100000000	10,84	11,97
		12,46	
		12,62	
mecA/mecC	10000000	15,41	16,02
		16,01	
		16,63	
mecA/mecC	1000000	19,63	18,96
		18,76	
		18,50	
mecA/mecC	100000	22,87	21,96
		21,34	
		21,67	
mecA/mecC	10000	26,63	26,65
		26,79	
		26,52	
mecA/mecC	1000	29,72	29,71
		29,91	
		29,50	
mecA/mecC	100	32,40	31,80
		32,19	
		30,81	
mecA/mecC	10	31,54	33,41
		34,07	
		34,61	
mecA/mecC	1,0	45,00	43,28
		45,00	
		39,83	

Sample	Concentration copies/ reaction	C _T -value	mean C _T
		ROX-channel	
PVL	100000000	10,55 10,08 10,00	10,21
PVL	10000000	13,71 13,67 13,67	13,68
PVL	1000000	16,94 17,03 16,82	16,93
PVL	100000	20,24 20,25 20,38	20,29
PVL	10000	23,68 23,69 23,72	23,70
PVL	1000	26,66 26,60 26,69	26,65
PVL	100	29,14 29,21 29,18	29,18
PVL	10	30,13 30,04 30,31	30,16
PVL	1,0	30,76 30,65 30,83	30,75

15.2 Linear Range

The linear range of the diarellaPVL-MRSA real time PCR was evaluated by analyzing logarithmic dilution series of synthetic DNA fragments.



15.3 Analytical Specificity

The specificity of the diarellaPVL-MRSA real time PCR was evaluated additionally with different other relevant viruses and bacteria found in clinical samples.

Results:

The diarellaPVL-MRSA real time PCR showed a positive result for the sample containing MRSA, whereas samples containing other pathogens were reliably tested negative. The results are shown in table 7 and table 8.

Table 7: Bacterial and viral pathogens tested for the determination of the analytical specificity of the diarellaPVL-MRSA real time PCR Kit.

Strain	FAM	ROX	Cy5	MRSA
Streptococcus agalactiae	negative	negative	negative	negative
Coxsackievirus Strain P.B.	negative	negative	negative	negative
Coxsackievirus Strain B.S.	negative	negative	negative	negative
Herpes simplex virus	negative	negative	negative	negative
Borrelia burgdorferi	negative	negative	negative	negative
S. uberis	negative	negative	negative	negative
Streptococcus dysagalactiae	negative	negative	negative	negative
Staphylococcus intermedius	negative	negative	negative	negative
Pseudomonas aeruginosa	negative	negative	negative	negative
Staphylococcus sciuri	negative	negative	negative	negative
Legionella pneumophila Serogroup 1	negative	negative	negative	negative

Table 8: Ring trial samples tested for the determination of the analytical specificity of the diarellaPVL-MRSA real time PCR Kit.

Sample	expected result	diarellaPVL-MRSA
1815391 cMRSA (S. aureus, oxaR, PVL-pos, spa:t 008)	positive	positive
1815392 MRSA (S. aureus, oxaR, PVL-neg)	positive	positive
1815393 Escherichia coli K12	negative	negative
1815394 cMSSA + CoNS (S. aureus, S. epidermidis oxaR, PVL-pos)	negative	negative
1715391 MRSA (S. aureus, oxaR, PVL-neg)	positive	positive
1715392 MSSA (SCCmec pos, mecA neg) (S. aureus, oxaS, PVL-neg, spa:t 310)	positive	positive
1715394 MRSA (S. aureus, oxaR, PVL-neg)	positive	positive
1615391 MSSA + CoNS (S. aureus, S. epidermidis oxaR, PVL-neg)	negative	negative
1615392 cMRSA (S. aureus, oxaR, PVL-pos, spa:t 310)	positive	positive
1615393 CoNS (S. epidermidis, oxaS)	negative	negative
1615394 cMRSA (S. aureus, oxaR, PVL-pos, spa:t 008)	positive	positive
1025391 MRSA + CoNS (S. aureus, CoNS, oxaR, PVL-neg)	positive	positive
1025392 MSSA + CoNS (S. aureus, CoNS, oxaS) "mecA dropout mutant"	positive	positive
1025393 cMRSA + CoNS (S. aureus, CoNS, oxaR, PVL-pos)	positive	positive
1025394 CoNS (oxaS)	negative	negative
1015391 MRSA ("Züricher Drogenstamm"; CHE 482) SCCmec PCRs (S. aureus, oxaR, PVL-neg)	positive	positive
1015393 MSSA + CoNS (S. aureus, CoNS, oxaR)	negative	negative
92901 cMSSA + CoNS (S. aureus, S. epidermidis oxaS, PVL-pos)	negative	negative
92902 cMRSA spa:t 310 (S. aureus, oxaR, PVL-pos)	positive	positive
92904 CoNS (S. epidermidis, oxaR)	negative	negative
91901 cMRSA spa:t 657 (S. aureus, oxaR, PVL-pos)	positive	positive
91902 MSSA + CoNS (S. epidermidis, oxaR)	negative	negative
1515391 MRSA SCCmec Typ V (S. aureus, oxaR, PVL-neg)	positive	positive
1515394 MSSA + CoNS (S. aureus, S. epidermidis oxaR, PVL-neg)	negative	negative
12-02 MRSA N315 (Core)	positive	positive
12-03 MRSA N315 (Core)	positive	positive
12-04 MSSA ATCC 29213 (Core)	negative	negative
12-05 MRSA ST398	positive	positive
12-06 MRSA N315	positive	positive
12-07 MRSA N315 (Core)	positive	positive
12-08 MSSA 29213 + MRCoNS 634 (Core)	negative	negative
12-09 MRSA N315 (Core)	positive	positive
12-10 MRSA Negativee MHB (Core)	negative	negative
12-11 MRSA "mecC"	positive	positive
12-12 E. coli ATCC 35218 (Core)	negative	negative

15.4 Precision

The precision of the diarellaPVL-MRSA real time PCR was determined as intra-assay variability, inter-assay variability and inter-lot variability.

Variability data are expressed by standard deviation and coefficient of variation. The data are based on quantification analyses of defined concentrations of MRSA SCCmec/orfX, MRSA mecA/mecC and PVL specific DNA and on the threshold cycle of the Control-DNA.

Table 9: Precision of diarellaPVL-MRSA real time PCR.















SCCmec/orfX	copies/rxn	Standard Deviation	Coefficient of Variation [%]
Intra-Assay Variability	1.000	0.43	1.49
Inter-Assay-Variability	1.000	0.30	1.06
Inter-Lot Variability	1.000	0.64	2.25

mecA/mecC	copies/rxn	Standard Deviation	Coefficient of Variation [%]
Intra-Assay Variability	1.000	1.22	4.20
Inter-Assay-Variability	1.000	0.26	0.87
Inter-Lot Variability	1.000	0.12	0.41

PVL	copies/rxn	Standard Deviation	Coefficient of Variation [%]
Intra-Assay Variability	100	0.10	0.36
Inter-Assay-Variability	100	0.15	0.52
Inter-Lot Variability	100	0.08	0.26

Control DNA	copies/rxn	Standard Deviation	Coefficient of Variation [%]
Intra-Assay Variability	100	0.71	2.44
Inter-Assay-Variability	100	0.22	0.74
Inter-Lot Variability	100	0.32	1.09

16 Abbreviations and Symbols

MRSA	Methicillin-resistant <i>Staphylococcus aureus</i>		Catalog number
MS-MRSA	Methicillin-suceptible MRSA, mecA dropout mutant		Contains sufficient for <n> test
SCCmec/orfX	Junction for <i>S. aureus</i> DNA and SCCmec cassette		Upper limit of temperature
MSSA	Methicillin-suceptible <i>Staphylococcus aureus</i>		Manufacturer
MR-ConS	Methicillin-resistant coagulase negative <i>Staphylococcus</i>		Use by YYYY-MM-DD
mecA / mecC	Two varaints of the methicillin resistance gene		Batch code
PVL	Panton-Valentine- Leucocidine		Content
DNA	Deoxyribonucleid Acid		Consult instructions for use
PCR	Polymerase Chain Reaction		<i>In vitro</i> diagnostic medical device
	Reaction Mix		European Conformity
	Positive Control		
	Negative Control		
	Control DNA		

17 Literature

- [1] Bundesgesundheitsbl 2014, 57, 696–732: Empfehlungen zur Prävention und Kontrolle von Methicillin-resistenten *Staphylococcus aureus*-Stämmen (MRSA) in medizinischen und pflegerischen Einrichtungen.
- [2] Centers for Disease Control and Prevention: Methicillin-resistant *Staphylococcus aureus*. www.cdc.gov/mrsa. May 16, 2016.
- [3] Epidemiologisches Bulletin, https://www.rki.de/DE/Content/Infekt/EpidBull/Archiv/2018/Ausgaben/05_18.pdf: Eigenschaften, Häufigkeit und Verbreitung von MRSA in Deutschland, update 2015/2016.